



THE SOUTH INDIA TEXTILE RESEARCH ASSOCIATION

CENTRE OF EXCELLENCE FOR MEDICAL TEXTILES

Physical, Chemical & Biological Testing Laboratories

(ISO/IEC 17025 NABL ACCREDITED)

13/37, Avinashi Road, Aerodrome Post, Coimbatore - 641 014, INDIA



Cert. Number:TC-6944

Care And Klein Essentials, Ref : Letter

ULR: TC694420600004292P

(Sample Tested at : R.H.65% +/- 2% and Temp. 21 Degree C +/- 1 Degree C)

| | |
|--|---|
| Differential Pressure - N95/N99 Mask** IS 16289:2014 Annexure C Pa/cm ² | P2005526-1 Described by the Customer : N 95 Face Mask 27.0 |
| PFET at 0.3 Microns - N95/N99 Mask ASTM F2299/F2299M-03 (2010) Percentage | P2005526-1 Described by the Customer : N 95 Face Mask 94.80 |

Following Disclaimer is required to be appear in the test reports of N 95/N99 respirator masks released by Microbiology and CoE-Physics labs.
Disclaimer: The test performed (as above) is specific to surgical face mask. SITRA does not have facility for evaluating the performance of N 95/N99 respirator masks. Necessary care may be taken by the end user while interpreting the above results for the intended use.

Particulate Filtration Efficiency Test Test Side : Face Side (Outside)

Area Tested : 33.17cm² Particle Size : 0.3µm

Laboratory Condition : 21°C, 50%RH

** Test(s) not covered under NABL Scope.

- End of Report -